



## EXHIBIT A


REVISIONS									
REV	DATE	BY	APP'D	DESCRIPTION	DATE	BY	APP'D	DESCRIPTION	DATE
1	11/1/88	JJC	WCP						
2	11/1/88	MJB	MJS						
3	11/1/88	MJB	MJS						
4	11/1/88	MJB	MJS						

[illegible]

# EXHIBIT A (cont.)




## EXHIBIT B

 <b>Medtronic</b>	<i>Neurological</i>	Document Number <b>288117-70205</b>	Rev/Version <b>1.0</b>	Sht 1 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report				

### Revision History:


Revision	Comments
1.0	Initial release for routing

# EXHIBIT B (cont.)

	<b>Medtronic</b>	<i>Neurological</i>	<b>Document Number</b> 288117-70205	<b>Rev/Version</b> 1.0	<b>Sht</b> 2 of 49
<b>Title: Neuro Patient Programmer Platform Electrical DVT Report</b>					

<b>1</b>	<b>INTRODUCTION .....</b>	<b>3</b>
1.1	Purpose.....	3
1.2	Scope .....	3
1.3	Document Overview .....	3
<b>2</b>	<b>REFERENCES AND DEFINITIONS .....</b>	<b>4</b>
2.1	Internal Medtronic References.....	4
	Note: Document revisions referenced in DVT Plan. ....	4
2.2	External References .....	4
2.3	Definitions, Acronyms, and Abbreviations .....	4
<b>3</b>	<b>TEST RESULTS SUMMARY .....</b>	<b>5</b>
3.1	Test Paths .....	6
<b>4</b>	<b>ELECTRICAL TESTS .....</b>	<b>7</b>
4.1	Power Source Tests .....	7
4.1.1	Current Drain Test .....	7
4.1.2	Supply Voltage Range Test.....	9
4.2	Input/Output Connections Tests .....	11
4.2.1	Keypad Interface Test.....	11
4.2.2	Display Interface Test .....	12
4.2.3	External Antenna Interface Test.....	14
4.2.4	Infrared Port Interface Test .....	16
4.2.5	Audio Transducer Test.....	17
4.2.6	Manufacturing/Test Interface Test.....	18
4.3	Internal Resources Tests .....	19
4.3.1	Memory Test .....	19
4.3.2	Real-Time Clock Backup Test.....	20
4.3.3	Real-Time Clock Accuracy Test.....	21
4.3.4	A/D Measurements Test .....	22
4.3.5	D/A Control Voltages Test.....	24
4.4	Transmit Telemetry (Downlink) Tests .....	25
4.4.1	Magnetic Field Intensity Test.....	25
4.4.2	Burst Characteristics Test.....	27
4.5	Receive Telemetry (Uplink) Tests .....	30
4.5.1	Detection Threshold Test.....	30
4.5.2	Detection Margin Test.....	32
4.5.3	Noise Immunity Test .....	34
4.5.4	Signal Distortion Test.....	37
4.5.5	Turnaround Time Test .....	40
4.5.6	Hold Drift Test.....	41
4.5.7	New-Battery FET Test.....	43
4.6	Telemetry Performance Tests.....	44
4.6.1	.....	44
<b>5</b>	<b>COMPLETION.....</b>	<b>49</b>

## EXHIBIT B (cont.)

 <b>Medtronic</b>	<i>Neurological</i>	<b>Document Number</b> <b>288117-70205</b>	<b>Rev/Version</b> 1.0	<b>Sht</b> 3 of 49
<b>Title: Neuro Patient Programmer Platform Electrical DVT Report</b>				

### 1 INTRODUCTION

This document is the electrical Design Verification Test (DVT) Report for the 37741 Patient Programmer Platform.

#### 1.1 Purpose

The purpose of this report is to document the results of test plan

#### 1.2 Scope


This report applies only to design verification testing of the 37741 Patient Programmer Platform.

#### 1.3 Document Overview

This document is organized as follows:

- Section 2 contains references and definitions.
- Section 3 contains a table with the list of tests, software revisions, sample sizes, and test results.
- Section 4 contains the results of the electrical design verification tests.

## EXHIBIT B (cont.)

	<b>Medtronic</b>	<i>Neurological</i>	<b>Document Number</b> <b>288117-70205</b>	<b>Rev/Version</b> <b>1.0</b>	<b>Sht</b> <b>4 of 49</b>
<b>Title: Neuro Patient Programmer Platform Electrical DVT Report</b>					

## 2 REFERENCES AND DEFINITIONS

This section identifies internal and external reference documents that augment the information provided in this document. It also defines terms, acronyms, and abbreviations used within the document.

### 2.1 Internal Medtronic References

Number	Name
120275	
215387	
288117-70040	
288117-70044	
288117-70029	
503011001	
288117-70200	

Note: Document revisions referenced in DVT Plan.

### 2.2 External References

*Reference the PEM Electrical Specification for external specification standards.*

### 2.3 Definitions, Acronyms, and Abbreviations

**ARB:** Arbitrary Waveform Generator

**ARB equipment:** One or more arbitrary waveform generators, used alone or in conjunction to generate sophisticated waveforms.

**DUT:** Device Under Test

**DVT:** Design Verification Test

**DVT Console:** The test console needed to perform the tests specified herein.

**ES:** Electrical Specification #120275


**GPB:** General Purpose Interface Bus

**PEM:** Patient Electronic Module

**PP:** Patient Programmer

**POR:** Power On Reset

## EXHIBIT B (cont.)

 <b>Medtronic</b>	<i>Neurological</i>	<b>Document Number</b> <b>288117-70205</b>	<b>Rev/Version</b> <b>1.0</b>	<b>Sht</b> <b>5 of 49</b>
<b>Title: Neuro Patient Programmer Platform Electrical DVT Report</b>				

### 3 Test Results Summary


Table 1 summarizes the results of all electrical design verification testing. Section 4 details each test setup, criteria, and results.

- Test data is stored as 288117-70200.
- Table 1 indicates test name, sample size, DUT software revision, Test Script Software revision, test path, and results.
- Test paths are shown in section 3.1.


Table 1

Test Name	DUT Software Revision	Script Software Test Revision	Test Path	Results
	22			PASS
	22			PASS
	22			PASS
	22			PASS
	22			PASS
	22			PASS
	22			PASS
	22			PASS
	22			PASS
	22			PASS
	22			PASS
	22			PASS
	22			PASS
	22			PASS
	22			PASS
	22			PASS
	22			PASS
	22			PASS
	22			PASS
	22			PASS
	22			PASS
	22			PASS
	22			PASS
	22			PASS
	22			PASS
	1			PASS

## EXHIBIT B (cont.)


 <b>Medtronic</b>	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 6 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report				

### 3.1 Test Paths





# EXHIBIT B (cont.)

	<b>Medtronic</b>	<i>Neurological</i>	<b>Document Number</b> 288117-70205	<b>Rev/Version</b> 1.0	<b>Sht</b> 7 of 49
<b>Title: Neuro Patient Programmer Platform Electrical DVT Report</b>					

## 4 ELECTRICAL TESTS

This section specifies electrical tests performed on the 37741 Patient Programmer Platform.

### 4.1 Power Source Tests

#### 4.1.1 Current Drain Test

##### 4.1.1.1 Objective

To verify the current drain meets the requirements specified in the *Power Source* section of the PEM Electrical Specification.

##### 4.1.1.2 Method and Equipment

##### 4.1.1.3 Test Cases

There are \_ test cases for transmit using all combinations of test values below:

Parameter	Test Values	Units

The

There are test cases using all combinations of test values below:

Parameter	Test Values	Units


There are test cases using two combinations of test values below:

Parameter	Test Values	Units

There are total test cases.

##### 4.1.1.4 Acceptance Criteria

## EXHIBIT B (cont.)

	<b>Medtronic</b>	<i>Neurological</i>	<b>Document Number</b> 288117-70205	<b>Rev/Version</b> 1.0	<b>Sht</b> 8 of 49
<b>Title: Neuro Patient Programmer Platform Electrical DVT Report</b>					

Operating Condition (Ref.)	Antenna	Duty Cycle (%)	Current Drain (mA) MAX		
			V	V	V
Row A	INT				
Row B	INT				
Row C	INT				
Row D	INT				
Row E	INT				
Row F	INT				
Row G	INT				
Row H	EXT				
Row I	INT				
Row J	INT				

Note 1:

4.1.1.5 Test Setup

- 1.
- 2.
- 3.
- 4.


4.1.1.6 Test Procedure

- 1.
- 2.

3.

4.

# EXHIBIT B (cont.)

	<b>Medtronic</b>	<i>Neurological</i>	<b>Document Number</b> 288117-70205	<b>Rev/Version</b> 1.0	<b>Sht</b> 9 of 49
<b>Title: Neuro Patient Programmer Platform Electrical DVT Report</b>					

## 4.1.1.7 RESULTS **PASS**

All devices met the acceptance criteria.

Operating Condition	Current Drain (mA) MAX																	
Row	Spec	Min	Max	Mean	Std Dev		Spec	Min	Max	Mean	Std Dev		Spec	Min	Max	Mean	Std Dev	
A																		
B																		
C																		
D																		
E																		
F																		
G																		
H																		
I																		
J																		

## 4.1.2 **Supply Voltage Range Test**


### 4.1.2.1 Objective

To verify the supply voltage range meets the requirements specified in the *Power Source* section of the PEM Electrical Specification.

### 4.1.2.2 Method and Equipment

### 4.1.2.3 Test Cases

# EXHIBIT B (cont.)

 <b>Medtronic</b>	<i>Neurological</i>	<b>Document Number</b> 288117-70205	<b>Rev/Version</b> 1.0	<b>Sht</b> 10 of 49
<b>Title: Neuro Patient Programmer Platform Electrical DVT Report</b>				

Parameter	Test Values	Units

The

There is test case without transmit:

Parameter	Test Values	Units

## 4.1.2.4 Acceptance Criteria

Operating Condition	Antenna	H-Bridge Drive Duty Cycle (%)	Min operating voltage (V)


## 4.1.2.5 Test Setup

- 1.
- 2.
- 3.
- 4.

## 4.1.2.6 Test Procedure

- 1.
- 2.
- 3.

## EXHIBIT B (cont.)

 <b>Medtronic</b>	<i>Neurological</i>	<b>Document Number</b> <b>288117-70205</b>	<b>Rev/Version</b> <b>1.0</b>	<b>Sht</b> <b>11 of 49</b>
<b>Title: Neuro Patient Programmer Platform Electrical DVT Report</b>				

**4.1.2.7 RESULTS PASS**

All devices met the acceptance criteria.

Operating Condition	Antenna	Supply Voltage Range (Volts)				
		Min	Max	Avg	Std Dev	

**4.2 Input/Output Connections Tests****4.2.1 Keypad Interface Test****4.2.1.1 Objective**

To verify the keypad interface meets the requirements specified in the *Input/Output Connections* section of the PEM Electrical Specification.

**4.2.1.2 Method and Equipment****4.2.1.3 Test Cases**


Parameter	Test Values	Units

**4.2.1.4 Acceptance Criteria****4.2.1.5 Test Setup**

- 1.
- 2.
- 3.

**4.2.1.6 Test Procedure**

# EXHIBIT B (cont.)

 <b>Medtronic</b>	<i>Neurological</i>	<b>Document Number</b> 288117-70205	<b>Rev/Version</b> 1.0	<b>Sht</b> 12 of 49
<b>Title: Neuro Patient Programmer Platform Electrical DVT Report</b>				

3.

## 4.2.1.7 RESULTS **PASS**

All devices met the acceptance criteria.

Tests	Keypad Interface (pass/fail)		
	Pass	Pass	Pass
	Pass	Pass	Pass

## 4.2.2 Display Interface Test

### 4.2.2.1 Objective

To verify the display interface meets the requirements specified in the *Input/Output Connections* section of the PEM Electrical Specification.


### 4.2.2.2 Method and Equipment

### 4.2.2.3 Test Cases

There are test cases using combinations of the test values below:

Parameter	Test Values	Units

# EXHIBIT B (cont.)

	<b>Medtronic</b>	<i>Neurological</i>	<b>Document Number</b> 288117-70205	<b>Rev/Version</b> 1.0	<b>Sht</b> 13 of 49
<b>Title: Neuro Patient Programmer Platform Electrical DVT Report</b>					

## 4.2.2.4 Acceptance Criteria

Test Parameters				Requirements			

## 4.2.2.5 Test Setup

- 1.
- 2.
- 3.

## 4.2.2.6 Test Procedure


- 1.
- 2.
- 3.
- 4.

## 4.2.2.7 RESULTS **PASS**

All devices met the acceptance criteria.

Test	Display Interface (pass/fail)		
	Pass	Pass	Pass
	Pass	Pass	Pass
	Pass	Pass	Pass

## EXHIBIT B (cont.)

	<b>Medtronic</b>	<i>Neurological</i>	<b>Document Number</b> <b>288117-70205</b>	<b>Rev/Version</b> <b>1.0</b>	<b>Sht</b> <b>14 of 49</b>
<b>Title: Neuro Patient Programmer Platform Electrical DVT Report</b>					

### 4.2.3 External Antenna Interface Test

#### 4.2.3.1 Objective

To verify the external antenna interface meets the requirements specified in the *Input/Output Connections* section of the PEM Electrical Specification.

#### 4.2.3.2 Method and Equipment

#### 4.2.3.3 Test Cases

There are test cases using all combinations of test values below:

Parameter	Test Values	Units

#### 4.2.3.4 Acceptance Criteria

- When an external antenna is connected, there should be no downlink from the internal antenna.
- When an external antenna is connected, the uP should detect that the antenna is connected.


External Antenna					
	Min	Max	Min	Max	Yes/No

#### 4.2.3.5 Test Setup

- 1.
- 2.
- 3.
- 4.
- 5.



EXHIBIT B (cont.)

 <b>Medtronic</b>	<i>Neurological</i>	Document Number <b>288117-70205</b>	Rev/Version 1.0	Sht 15 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report				

4.2.3.6 Test Procedure

1.

2.


3.

4.

4.2.3.7 RESULTS **PASS**

All devices met the acceptance criteria.

# EXHIBIT B (cont.)

 <b>Medtronic</b>	<i>Neurological</i>	<b>Document Number</b> 288117-70205	<b>Rev/Version</b> 1.0	<b>Sht</b> 16 of 49
<b>Title: Neuro Patient Programmer Platform Electrical DVT Report</b>				

Test Configuration	Test
	A
	B

Test	External Antenna Interface (A/m)													
	Min	Max	Mean	Std dev		Min	Max	Mean	Std dev		Min	Max	Mean	Std dev
A														
B														

## 4.2.4 Infrared Port Interface Test

### 4.2.4.1 Objective

To verify the infrared port interface meets the requirements specified in the *Input/Output Connections* section of the PEM Electrical Specification. [PTPROG\_PEMT-0006:1]

### 4.2.4.2 Method and Equipment

### 4.2.4.3 Test Cases

There are test cases using all combinations of test values below:

Parameter	Test Values	Units


### 4.2.4.4 Acceptance Criteria

All	All	None

### 4.2.4.5 Test Setup

- 1.
- 2.

## EXHIBIT B (cont.)

	<b>Medtronic</b>	<i>Neurological</i>	<b>Document Number</b> 288117-70205	<b>Rev/Version</b> 1.0	<b>Sht</b> 17 of 49
<b>Title: Neuro Patient Programmer Platform Electrical DVT Report</b>					

3.

### 4.2.4.6 Test Procedure

1.

2.

3.

4.

### 4.2.4.7 RESULTS **PASS**

All devices met the acceptance criteria.

Voltage (V)	Infrared (pass/fail)								
	Pass	Pass	Pass	Pass	Pass	Pass	Pass	Pass	Pass
	Pass	Pass	Pass	Pass	Pass	Pass	Pass	Pass	Pass
	Pass	Pass	Pass	Pass	Pass	Pass	Pass	Pass	Pass

## 4.2.5 Audio Transducer Test

### 4.2.5.1 Objective


To verify the audio transducer meets the requirements specified in the *Input/Output Connections* section of the PEM Electrical Specification.

### 4.2.5.2 Method and Equipment

### 4.2.5.3 Test Cases

There are test cases using all combinations of test values below:

## EXHIBIT B (cont.)

	<b>Medtronic</b>	<i>Neurological</i>	<b>Document Number</b> 288117-70205	<b>Rev/Version</b> 1.0	<b>Sht</b> 18 of 49
<b>Title: Neuro Patient Programmer Platform Electrical DVT Report</b>					

Parameter	Test Values	Units

4.2.5.4 Acceptance Criteria


4.2.5.5 Test Setup

- 1.
- 2.
- 3.
- 4.
- 5.

4.2.5.6 Test Procedure

- 1.
- 2.
- 3.
- 4.

4.2.5.7 RESULTS **PASS**


All devices met the acceptance criteria.

Audio Transducer (dB SPL)														
Min	Max	Mean	Std dev		Min	Max	Mean	Std dev		Min	Max	Mean	Std dev	

4.2.6 **Manufacturing/Test Interface Test**

Manufacturing requirements defined in Test Specification, Patient Programmer, 215387.

# EXHIBIT B (cont.)

	<b>Medtronic</b>	<i>Neurological</i>	<b>Document Number</b> 288117-70205	<b>Rev/Version</b> 1.0	<b>Sht</b> 19 of 49
<b>Title: Neuro Patient Programmer Platform Electrical DVT Report</b>					

## 4.3 Internal Resources Tests

### 4.3.1 Memory Test

#### 4.3.1.1 Objective

To verify the internal memory resources meet the requirements specified in the *Internal Resources* section of the PEM Electrical Specification.

#### 4.3.1.2 Method and Equipment

#### 4.3.1.3 Test Cases

There are test cases using all combinations of test values below:

Parameter	Test Values	Units

#### 4.3.1.4 Acceptance Criteria

All	Pass

#### 4.3.1.5 Test Setup

- 1.
- 2.
- 3.

#### 4.3.1.6 Test Procedure


- 1.
- 2.
- 3.
- 4.

#### 4.3.1.7 RESULTS **PASS**

All devices met the acceptance criteria.

Test	Memory (pass/fail)		
	Pass	Pass	Pass
	Pass	Pass	Pass
	Pass	Pass	Pass

## EXHIBIT B (cont.)

	<b>Medtronic</b>	<i>Neurological</i>	<b>Document Number</b> 288117-70205	<b>Rev/Version</b> 1.0	<b>Sht</b> 20 of 49
<b>Title: Neuro Patient Programmer Platform Electrical DVT Report</b>					

### 4.3.2 Real-Time Clock Backup Test

#### 4.3.2.1 Objective

To verify the real-time clock backup meets the requirements specified in the *Internal Resources* section of the PEM Electrical Specification.

#### 4.3.2.2 Method and Equipment

#### 4.3.2.3 Test Cases

There is one test case below:

Parameter	Test Value	Units

#### 4.3.2.4 Acceptance Criteria

Test Case	Min Time w/o power (min)

#### 4.3.2.5 Test Setup

- 1.
- 2.
- 3.


#### 4.3.2.6 Test Procedure

- 1.
- 2.
- 3.
- 4.
- 5.

#### 4.3.2.7 RESULTS **PASS**

All devices met the acceptance criteria.

## EXHIBIT B (cont.)

	<b>Medtronic</b>	<i>Neurological</i>	<b>Document Number</b> 288117-70205	<b>Rev/Version</b> 1.0	<b>Sht</b> 21 of 49
<b>Title: Neuro Patient Programmer Platform Electrical DVT Report</b>					

Test	Real-Time Backup (pass/fail)		
	Pass	Pass	Pass

### 4.3.3 Real-Time Clock Accuracy Test

#### 4.3.3.1 Objective

To verify the real-time clock accuracy meets the requirements specified in the *Internal Resources* section of the PEM Electrical Specification.


#### 4.3.3.2 Method and Equipment

#### 4.3.3.3 Test Cases

There are test cases (actually measurement points) using all combinations of test values below:

Parameter	Test Value	Units

#### 4.3.3.4 Acceptance Criteria



#### 4.3.3.5 Test Setup

- 1.
- 2.


#### 4.3.3.6 Test Procedure

- 1.
- 2.

#### 4.3.3.7 RESULTS **PASS**

All devices met the acceptance criteria.

# EXHIBIT B (cont.)

 <b>Medtronic</b>	<i>Neurological</i>	<b>Document Number</b> 288117-70205	<b>Rev/Version</b> 1.0	Sht 22 of 49
<b>Title: Neuro Patient Programmer Platform Electrical DVT Report</b>				

	Real Time Clock Accuracy (seconds)					

## 4.3.4 A/D Measurements Test

### 4.3.4.1 Objective

To verify the A/D measurement accuracy meets the requirements specified in the *Internal Resources* section of the PEM Electrical Specification.

### 4.3.4.2 Method and Equipment

### 4.3.4.3 Test Cases

There are      test cases using the test values below:

Parameter	Test Values	Units

### 4.3.4.4 Acceptance Criteria


A/D Voltage	Test Value	Max Error (%)

### 4.3.4.5 Test Setup

- 1.
- 2.
- 3.
- 4.



EXHIBIT B (cont.)

 <b>Medtronic</b>	<i>Neurological</i>	<b>Document Number</b> 288117-70205	<b>Rev/Version</b> 1.0	<b>Sht</b> 23 of 49
<b>Title: Neuro Patient Programmer Platform Electrical DVT Report</b>				

5.

4.3.4.6 Test Procedure

1.

2.


3.

4.

4.3.4.7 RESULTS **PASS**

All devices met the acceptance criteria.

## EXHIBIT B (cont.)

 <b>Medtronic</b>	<i>Neurological</i>	<b>Document Number</b> <b>288117-70205</b>	<b>Rev/Version</b> <b>1.0</b>	<b>Sht</b> <b>24 of 49</b>
<b>Title: Neuro Patient Programmer Platform Electrical DVT Report</b>				

[illegible]

#### 4.3.5 D/A Control Voltages Test

#### 4.3.5.1 Objective


To verify the D/A accuracy meets the requirements specified in the *Internal Resources* section of the PEM Electrical Specification.

#### 4.3.5.2 Method and Equipment

#### 4.3.5.3 Test Cases

There are test cases using all combinations of test values below:

## EXHIBIT B (cont.)

 <b>Medtronic</b>	<i>Neurological</i>	<b>Document Number</b> <b>288117-70205</b>	<b>Rev/Version</b> <b>1.0</b>	<b>Sht</b> <b>25 of 49</b>
<b>Title: Neuro Patient Programmer Platform Electrical DVT Report</b>				

Parameter	Test Value	Units

#### 4.3.5.4 Acceptance Criteria

D/A Voltage	Measurement point	Max % Error

#### 4.3.5.5 Test Setup

- 1.
- 2.
- 3.
- 4.

#### 4.3.5.6 Test Procedure

- 1.
- 2.
- 3.

4.3.5.7 RESULTS PASS

All devices met the acceptance criteria.

[illegible]


#### 4.4 Transmit Telemetry (Downlink) Tests

#### 4.4.1 Magnetic Field Intensity Test

#### 4.4.1.1 Objective

To verify downlink magnetic field intensity meets the requirements specified in the *Transmit Telemetry (Downlink)* section of the PEM Electrical Specification.

## EXHIBIT B (cont.)

 <b>Medtronic</b>	<i>Neurological</i>	<b>Document Number</b> <b>288117-70205</b>	<b>Rev/Version</b> <b>1.0</b>	<b>Sht</b> <b>26 of 49</b>
<b>Title: Neuro Patient Programmer Platform Electrical DVT Report</b>				

#### 4.4.1.2 Method and Equipment

#### 4.4.1.3 Test Cases

There are test cases at kHz using all combinations of test values below:

Category	Time 1	Time 2	Time 3	Time 4
A	10	15	25	45
B	10	15	20	35
C	10	15	20	25
D	10	15	15	15
E	10	15	15	10

#### 4.4.1.4 Acceptance Criteria


#### 4.4.1.5 Test Setup

- 1.
- 2.
- 3.
- 4.
- 5.

#### 4.4.1.6 Test Procedure

- 1.
- 2.
- 3.
- 4.

## EXHIBIT B (cont.)

 <b>Medtronic</b>	<i>Neurological</i>	<b>Document Number</b> <b>288117-70205</b>	<b>Rev/Version</b> <b>1.0</b>	<b>Sht</b> <b>27 of 49</b>
<b>Title: Neuro Patient Programmer Platform Electrical DVT Report</b>				

5.

**4.4.1.7 RESULTS PASS**

All devices met the acceptance criteria.

	Magnetic Field Intensity (A/m)

**4.4.2 Burst Characteristics Test****4.4.2.1 Objective**


To verify downlink burst characteristics of width, rise time, fall time, frequency, and overshoot meet the requirements specified in the *Transmit Telemetry (Downlink)* section of the PEM Electrical Specification.

**4.4.2.2 Method and Equipment****4.4.2.3 Test Cases**

There are test cases using all combinations of test values below:

Parameter	Test Values	Units

EXHIBIT B (cont.)

 <b>Medtronic</b>	<i>Neurological</i>	<b>Document Number</b> 288117-70205	<b>Rev/Version</b> 1.0	<b>Sht</b> 28 of 49
<b>Title: Neuro Patient Programmer Platform Electrical DVT Report</b>				

4.4.2.4 Acceptance Criteria

--

4.4.2.5 Test Setup

- 1.
- 2.
- 3.
- 4.
- 5.

4.4.2.6 Test Procedure

- 1.
- 2.
- 3.
- 4.
- 5.

4.4.2.7 RESULTS **PASS**

All devices met the acceptance criteria.


### Neurological

Rev/Version  
1.0

**Title: Neuro Patient Programmer Platform Electrical DVT Report**

[illegible]

# EXHIBIT B (cont.)

 <b>Medtronic</b>	<i>Neurological</i>	<b>Document Number</b> <b>288117-70205</b>	<b>Rev/Version</b> <b>1.0</b>	<b>Sht</b> <b>30 of 49</b>
<b>Title: Neuro Patient Programmer Platform Electrical DVT Report</b>				

## 4.5 Receive Telemetry (Uplink) Tests


### 4.5.1 Detection Threshold Test

#### 4.5.1.1 Objective

To verify uplink detection threshold (i.e. receiver sensitivity) meets the requirements specified in the *Receive Telemetry (Uplink)* section of the PEM Electrical Specification.

#### 4.5.1.2 Method and Equipment

**Figure 1: Example Uplink Detection Threshold Test Waveforms**



#### 4.5.1.3 Test Cases

There are test cases using all combinations of test values below:

Parameter	Test Values	Units


The supply voltage is 2.5 V.



Sht  
31 of 49

[illegible]

## EXHIBIT B (cont.)

 <b>Medtronic</b>	<i>Neurological</i>	<b>Document Number</b> <b>288117-70205</b>	<b>Rev/Version</b> <b>1.0</b>	<b>Sht</b> <b>32 of 49</b>
<b>Title: Neuro Patient Programmer Platform Electrical DVT Report</b>				

Antenna	Telemetry	Maximum Input Level (pass/fail)		
		Pass	Pass	Pass
		Pass	Pass	Pass
		Pass	Pass	Pass
		Pass	Pass	Pass
		Pass	Pass	Pass
		Pass	Pass	Pass


### 4.5.2 Detection Margin Test

#### 4.5.2.1 Objective

To verify uplink detection margin meets the requirements specified in the *Receive Telemetry (Uplink)* section of the PEM Electrical Specification.

#### 4.5.2.2 Method and Equipment


**Figure 2: Example Uplink Detection Margin Test Waveforms**



#### 4.5.2.3 Test Cases

There are     test cases using all combinations of test values below:

# EXHIBIT B (cont.)

	<b>Medtronic</b>	<i>Neurological</i>	<b>Document Number</b> 288117-70205	<b>Rev/Version</b> 1.0	<b>Sht</b> 33 of 49
<b>Title: Neuro Patient Programmer Platform Electrical DVT Report</b>					

Parameter	Test Values	Units

## 4.5.2.4 Acceptance Criteria

Telemetry Type	Data Bursts	Amplitude A1	Antenna	Detection Margin (Uplink dB)	
				Min	Max

## 4.5.2.5 Test Setup

- 1.
- 2.
- 3.
- 4.
- 5.


## 4.5.2.6 Test Procedure

- 1.
- 2.
- 3.
- 4.

## 4.5.2.7 RESULTS **PASS**

All devices met the acceptance criteria.

## EXHIBIT B (cont.)

 <b>Medtronic</b>	<i>Neurological</i>	<b>Document Number</b> 288117-70205	<b>Rev/Version</b> 1.0	<b>Sht</b> 34 of 49
<b>Title: Neuro Patient Programmer Platform Electrical DVT Report</b>				

[illegible]


### 4.5.3 Noise Immunity Test

#### 4.5.3.1 Objective


To verify uplink noise immunity meets the requirements specified in the *Receive Telemetry (Uplink)* section of the PEM Electrical Specification.

#### 4.5.3.2 Method and Equipment

# EXHIBIT B (cont.)

	<b>Medtronic</b>	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 35 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report					

**Figure 3: Example Uplink Noise Immunity Test Waveforms**




## 4.5.3.3 Test Cases

There are      test cases using all combinations of test values below:

Parameter	Test Values	Units

## EXHIBIT B (cont.)

 <b>Medtronic</b>	<i>Neurological</i>	<b>Document Number</b> 288117-70205	<b>Rev/Version</b> 1.0	Sht 36 of 49
<b>Title: Neuro Patient Programmer Platform Electrical DVT Report</b>				

### 4.5.3.4 Acceptance Criteria

Telemetry Type	Uplink Level A1 (dB)	Antenna	Min E-Noise Immunity (dB)	Min H-Noise Immunity (dB)

### 4.5.3.5 Test Setup

- 1.
- 2.
- 3.
- 4.
- 5.
- 6.


### 4.5.3.6 Test Procedure

- 1.
- 2.
- 3.
- 4.
- 5.

### 4.5.3.7 RESULTS **PASS**

All devices met the acceptance criteria.

## EXHIBIT B (cont.)

 <b>Medtronic</b>	<i>Neurological</i>	<b>Document Number</b> <b>288117-70205</b>	<b>Rev/Version</b> <b>1.0</b>	<b>Sht</b> <b>37 of 49</b>
<b>Title: Neuro Patient Programmer Platform Electrical DVT Report</b>				

[illegible]


#### 4.5.4 Signal Distortion Test

#### 4.5.4.1 Objective

To verify uplink signal distortion meets the requirements specified in the *Receive Telemetry (Uplink)* section of the PEM Electrical Specification.

#### 4.5.4.2 *Method and Equipment*

## EXHIBIT B (cont.)

	<b>Medtronic</b>	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 38 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report					

### 4.5.4.3 Test Cases

Parameter	Test Values	Units

There are    test cases for Tel A, and    test cases for Tel N.

### 4.5.4.4 Acceptance Criteria

Telemetry Type	Uplink Level A1 (dB)	Antenna	Interval Distortion (μS )	Active/Idle Distortion (μS )

### 4.5.4.5 Test Setup

- 1.
- 2.
- 3.
- 4.
- 5.

### 4.5.4.6 Test Procedure


- 1.
- 2.
- 3.
- 4.

### 4.5.4.7 RESULTS **PASS**

All devices met the acceptance criteria.



## EXHIBIT B (cont.)

 <b>Medtronic</b>	<i>Neurological</i>	<b>Document Number</b> <b>288117-70205</b>	<b>Rev/Version</b> <b>1.0</b>	<b>Sht</b> <b>39 of 49</b>
<b>Title: Neuro Patient Programmer Platform Electrical DVT Report</b>				

[illegible][illegible]

### Neurological

**288117-70205**


Sht  
40 of 49

[illegible]

There are test cases using all combinations of test values below:

Parameter	Test Values	Units

## EXHIBIT B (cont.)

	<b>Medtronic</b>	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 41 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report					

### 4.5.5.4 Acceptance Criteria

Supply Voltage	H-Bridge Drive Duty Cycle	Turnaround Time (mS )

### 4.5.5.5 Test Setup

- 1.
- 2.
- 3.

### 4.5.5.6 Test Procedure

- 1.
- 2.
- 3.
- 4.

### 4.5.5.7 RESULTS **PASS**

All devices met the acceptance criteria.

	Turnaround Time (pass/fail)
Test	


### 4.5.6 **Hold Drift Test**

#### 4.5.6.1 Objective

To verify the hold drift meets the requirements specified in the *Receive Telemetry (Uplink)* section of the PEM Electrical Specification.

#### 4.5.6.2 Method and Equipment

# EXHIBIT B (cont.)

 <b>Medtronic</b>	<i>Neurological</i>	<b>Document Number</b> 288117-70205	<b>Rev/Version</b> 1.0	<b>Sht</b> 42 of 49
<b>Title: Neuro Patient Programmer Platform Electrical DVT Report</b>				

## 4.5.6.3 Test Cases

There is test case:

Parameter	Uplink Level	Units

## 4.5.6.4 Acceptance Criteria

Time after hold circuit enabled	Max Hold Drift


## 4.5.6.5 Test Setup

- 1.
- 2.
- 3.

## 4.5.6.6 Test Procedure


- 1.
- 2.
- 3.
- 4.
- 5.
- 6.
- 7.
- 8.
- 9.
- 10

# EXHIBIT B (cont.)

 <b>Medtronic</b>	<i>Neurological</i>	<b>Document Number</b> <b>288117-70205</b>	<b>Rev/Version</b> <b>1.0</b>	<b>Sht</b> <b>43 of 49</b>
<b>Title: Neuro Patient Programmer Platform Electrical DVT Report</b>				

## 4.5.6.7 RESULTS **PASS**

All devices met the acceptance criteria.

Hold Drift (mV)


## 4.5.7 New-Battery FET Test

### 4.5.7.1 Objective

To verify that enabling the new-battery FET circuit reduces the receiver noise floor (ambient RF energy detected by the receiver circuit) when new batteries are used.

### 4.5.7.2 Method and Equipment

### 4.5.7.3 Test Cases

There is test case:

Parameter	Uplink Level	Units


### 4.5.7.4 Acceptance Criteria

Supply Voltage	New-Battery FET	RSSI Peak

### 4.5.7.5 Test Setup

- 1.
- 2.
- 3.

## EXHIBIT B (cont.)

 <b>Medtronic</b>	<i>Neurological</i>	<b>Document Number</b> <b>288117-70205</b>	<b>Rev/Version</b> <b>1.0</b>	<b>Sht</b> <b>44 of 49</b>
<b>Title: Neuro Patient Programmer Platform Electrical DVT Report</b>				

### 4.5.7.6 Test Procedure

- 1.
- 2.
- 3.
- 4.
- 5.
- 6.
- 7.

### 4.5.7.7 RESULTS **PASS**

All devices met the acceptance criteria.

New-Battery FET (mV)													
Min	Max	Mean	Std dev		Min	Max	Mean	Std dev		Min	Max	Mean	Std dev

## 4.6 Telemetry Performance Tests


### 4.6.1 Telemetry Map Area at a Fixed Distance Test

#### 4.6.1.1 Objective

To verify telemetry performance in terms of map area at a fixed distance meets the requirements specified in the *Telemetry Performance* section of the PEM Electrical Specification.

#### 4.6.1.2 Method and Equipment

## EXHIBIT B (cont.)

 <b>Medtronic</b>	<i>Neurological</i>	<b>Document Number</b> 288117-70205	<b>Rev/Version</b> 1.0	<b>Sht</b> 45 of 49
<b>Title: Neuro Patient Programmer Platform Electrical DVT Report</b>				

4.6.1.3 Test Cases

Parameter	Test Values	Units

There are    test cases.

4.6.1.4 Acceptance Criteria

IPG	Antenna	Map Area @ 5cm

4.6.1.5 Test Setup


- 1.
- 2.

4.6.1.6 Test Procedure


- 1.
- 2.
- 3.
- 4.
- 5.
- 6.
- 7.
- 8.

4.6.1.7 RESULTS **PASS**

# EXHIBIT B (cont.)

	<b>Medtronic</b>	<i>Neurological</i>	<b>Document Number</b> 288117-70205	<b>Rev/Version</b> 1.0	<b>Sht</b> 46 of 49
<b>Title: Neuro Patient Programmer Platform Electrical DVT Report</b>					

## 4.6.1.7.1 Internal Antenna Map @



## 4.6.1.7.2 Internal Antenna @









EXHIBIT B (cont.)

 <b>Medtronic</b>	<i>Neurological</i>	<b>Document Number</b> 288117-70205	<b>Rev/Version</b> 1.0	<b>Sht</b> 47 of 49
<b>Title: Neuro Patient Programmer Platform Electrical DVT Report</b>				

4.6.1.7.3 External Antenna Map @




4.6.1.7.4 External Antenna @




4.6.1.7.5 Photo of test fixture showing

EXHIBIT B (cont.)

 <b>Medtronic</b>	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 48 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report				

in this photo.



4.6.1.7.6 Photo of





EXHIBIT B (cont.)

 <b>Medtronic</b>	<i>Neurological</i>	<b>Document Number</b> <b>288117-70205</b>	<b>Rev/Version</b> <b>1.0</b>	<b>Sht</b> <b>49 of 49</b>
<b>Title: Neuro Patient Programmer Platform Electrical DVT Report</b>				

**5 COMPLETION**

This paragraph concludes this test specification.

wdkSpace - Microsoft Internet Explorer

File Edit View Favorites Tools Help

Back Forward Stop Home Reload

Address <http://mspreg03.corp.medtronic.com:7111/npd30/wdkSpace/wdkSpace.jsp>

User: Alex Toy Docbase: npd

Docbases Inbox Workflow Status Search Checked-out

288117-70205  
Word 97 / 2000 document

Version	Last Modified	Lock Owner
Approved, CURRENT, 2.0	06/28/2003	
1.0	10/07/2002	

-Select action- -Select action-

OK

Local Intranet

# Test Path #1 from DVT Plan 288117-70020 Section 7.0

DVT Pre-Test Performed to verify operational units.

Buttons

Serial Number	operational	Audio	LCD	Battery contact	Battery Door	Real time clock	IR	Backlight	Communication	Results
NJD000018P	x	x	x	x	x	x	x	x	x	OK
NJD000019P	x	x	x	x	x	x	x	x	x	OK
NJD000020P	x	x	x	x	x	x	x	x	x	OK
NJD000022P	x	x	x	x	x	x	x	x	x	OK
NJD000024P	x	x	x	x	x	x	x	x	x	OK
NJD000025P	x	x	x	x	x	x	x	x	x	OK
NJD000026P	x	x	x	x	x	x	x	x	x	OK
NJD000028P	x	x	x	x	x	x	x	x	x	OK
NJD000031P	x	x	x	x	x	x	x	x	x	OK
NJD000033P	x	x	x	x	x	x	x	x	x	OK
NJD000034P	x	x	x	x	x	x	x	x	x	OK
NJD000035P	x	x	x	x	x	x	x	x	x	OK
NJD000036P	x	x	x	x	x	x	x	x	x	OK
NJD000037P	x	x	x	x	x	x	x	x	x	OK
NJD000077P	x	x	x	x	x	x	x	x	x	OK
NJD000078P	x	x	x	x	x	x	x	x	x	OK
NJD000079P	x	x	x	x	x	x	x	x	x	OK
NJD000080P	x	x	x	x	x	x	x	x	x	OK
NJD000138P	x	x	x	x	x	x	x	x	x	OK
NJD000139P	x	x	x	x	x	x	x	x	x	OK
NJD000140P	x	x	x	x	x	x	x	x	x	OK
NJD000149P	x	x	x	x	x	x	x	x	x	OK

Testing performed by

Date:

23-May-02

EQUIPMENT:

Exhibit D (cont.)

PAR# 5365      TEST PLAN: 288117-70020      SUMMERY SHEET      TECH:

Patient Programmer for Neuro devices.

DATE: 29 MAY 02      INITIAL VISUAL & ELECTRICAL

SERIAL#	VISUAL	Requestor
NJD000018P	O.K.	did functional X
NJD000019P	O.K.	X
NJD000020P	O.K.	X
NJD000022P	O.K.	X
NJD000024P	O.K.	X
NJD000025P	O.K.	X
NJD000026P	O.K.	X
NJD000028P	O.K.	X
NJD000031P	O.K.	X
NJD000033P	O.K.	X
NJD000034P	O.K.	X
NJD000035P	O.K.	X
NJD000036P	O.K.	X
NJD000037P	O.K.	X
NJD000077P	O.K.	X
NJD000078P	O.K.	X
NJD000079P	O.K.	X
NJD000080P	O.K.	X
NJD000138P	O.K.	X
NJD000139P	O.K.	X
NJD000140P	O.K.	X
NJD000149P	O.K.	X
	3-Jun	

RESULTS: NO ANOMALIES NOTED

SUMMARY SHEET

TEST PLAN: 288117-70020

Patient Programmer for Neuro devices. Life cycle of battery contacts and door, and external antenna jack.

DATE: 19-Jun-02

Subject samples

Serial Number	Battery Door			Battery External Contact Antenna			Dimension			Weight						
	6.3.3	cycles		6.3.4	cycles		6.3.1	Length	Width	Ht.	6.3.2	oz.	2 AA batteries	oz.	Total Weight	oz.
288117-70020 test number																

288117-70020 test

Serial Number	6.3.3	cycles		6.3.4	cycles		6.3.5	cycles
NJD000018P								
NJD000019P								
NJD000020P								
NJD000022P								
NJD000024P								
NJD000025P								
NJD000026P								
NJD000028P								
NJD000031P								
NJD000033P								
NJD000034P								
NJD000035P								
NJD000036P								
NJD000037P								
NJD000077P								
NJD000078P								
NJD000079P								
NJD000080P								
NJD000138P								
NJD000139P								
NJD000140P								
NJD000149P								

Average

Exhibit D (cont.)

Test Path #1

DVT Test Data for 288117-70020

Revision 4.0

Std Dev  
Dimensions per print 502814

EQUIPMENT:



PAR# 5365

TEST PLAN: 288117-70020

SUMMARY SHEET

TECH:

Patient Programmer for Neuro devices.      Storage Temperature paragraph 6.2.2 of test plan.

DATE:      19-Jun-02    All Functional Testing done per 6.1 except backlight and IR port.

Subject samples to low temp. storage of    degrees F for    hours then    degrees F for    hours.

Functional test samples post each temperature storage.

Serial #	Functional	Functional
NJD000018P		
NJD000019P		
NJD000020P		
NJD000022P		
NJD000024P		
NJD000025P		
NJD000026P		
NJD000028P		
NJD000031P		
NJD000033P		
NJD000034P		
NJD000035P		
NJD000036P		
NJD000037P		
NJD000077P		
NJD000078P		
NJD000079P		
NJD000080P		
NJD000138P		
NJD000139P		
NJD000140P		
NJD000149P		
Date: Complete	18-Jun	18-Jun    19-Jun    19-Jun

NOTES:    A=  
             B=  
             C=

Results:

Test Path #1

DVT Test Data for 288117-70020

Revision 4.0

EQUIPMENT:

PAR# 5365      TEST PLAN: 288117-70020      TECH:  
 Patient Programmer for Neuro devices.      Operating Temperature paragraph 6.2.1 of test plan.  
 DATE:      4-Jun-02      All Functional Testing done per 6.1 except backlight and IR port.  
 Subject samples to Low temp. storage of      degrees F for      hours then      degrees F for      hours.

Serial #	Low temp.	Functional	High Temp.	Functional
NJD000018P				
NJD000019P				
NJD000020P				
NJD000022P				
NJD000024P				
NJD000025P				
NJD000026P				
NJD000028P				
NJD000031P				
NJD000033P				
NJD000034P				
NJD000035P				
NJD000036P				
NJD000037P				
NJD000077P				
NJD000078P				
NJD000079P				
NJD000080P				
NJD000138P				
NJD000139P				
NJD000140P				
NJD000149P				
Date: Complete	4-Jun	4-Jun	5-Jun	5-Jun
NOTES: A=				

Results:

EQUIPMENT:

PAR# 5365

TEST PLAN: 288117-70020

SUMMARY SHEET

TECH:

Patient Programmer for Neuro devices.

DATE: 20-Jun-02 Thermal Shock paragraph 6.2.3 of test plan.

Subject samples to cycles of degrees F, 1 degrees F, then 1  
Dwell at each temperature for 1 hour. All Functional Testing done per 6.1 except backlight and IR port.

Serial #	Thermal Shock		Functional Testing		Visual
NJD000018P					
NJD000019P					
NJD000020P					
NJD000022P					
NJD000024P					
NJD000025P					
NJD000026P					
NJD000028P					
NJD000031P					
NJD000033P					
NJD000034P					
NJD000035P					
NJD000036P					
NJD000037P					
NJD000077P					
NJD000078P					
NJD000079P					
NJD000080P					
NJD000138P					
NJD000139P					
NJD000140P					
NJD000149P					

NOTES: A=

RESULTS:

EQUIPMENT:

PAR# 5365

TEST PLAN: 288117-70020

Patient Programmer for Neuro devices.

DATE: 21-Jun-02

Chemical Resistance paragraph 6.2.7 of test plan.

Subject samples to

TECH:

SUMMARY SHEET

Serial #	Chemical Testing	Visual
NJD000018P		
NJD000019P		
NJD000020P		
NJD000022P		
NJD000024P		
NJD000025P		
NJD000026P		
NJD000028P		
NJD000031P		
NJD000033P		
NJD000034P		
NJD000035P		
NJD000036P		
NJD000037P		
NJD000077P		
NJD000078P		
NJD000079P		
NJD000080P		
NJD000138P		
NJD000139P		
NJD000140P		
NJD000149P		

RESULTS:

EQUIPMENT:

# Test Path #2 from DVT Plan 288117-70020 Section 7.0

DVT Pre-Test Performed to verify operational units.

Serial Number	Buttons		Audio	LCD	Battery		Door	Real time clock	IR	Backlight	Communication	
	operational	operational			contact	contact					n	Results
NJD000109P	x		x	x	x		x	x	x	x	x	OK
NJD000110P	x		x	x	x		x	x	x	x	x	OK
NJD000111P	x		x	x	x		x	x	x	x	x	OK
NJD000113P	x		x	x	x		x	x	x	x	x	OK
NJD000114P	x		x	x	x		x	x	x	x	x	OK
NJD000116P	x		x	x	x		x	x	x	x	x	OK
NJD000119P	x		x	x	x		x	x	x	x	x	OK
NJD000120P	x		x	x	x		x	x	x	x	x	OK
NJD000121P	x		x	x	x		x	x	x	x	x	OK
NJD000122P	x		x	x	x		x	x	x	x	x	OK
NJD000123P	x		x	x	x		x	x	x	x	x	OK
NJD000124P	x		x	x	x		x	x	x	x	x	OK
NJD000126P	x		x	x	x		x	x	x	x	x	OK
NJD000127P	x		x	x	x		x	x	x	x	x	OK
NJD000128P	x		x	x	x		x	x	x	x	x	OK
NJD000129P	x		x	x	x		x	x	x	x	x	OK
NJD000130P	x		x	x	x		x	x	x	x	x	OK
NJD000131P	x		x	x	x		x	x	x	x	x	OK
NJD000133P	x		x	x	x		x	x	x	x	x	OK
NJD000134P	x		x	x	x		x	x	x	x	x	OK
NJD000136P	x		x	x	x		x	x	x	x	x	OK
NJD000137P	x		x	x	x		x	x	x	x	x	OK

Exhibit D (cont.)

Testing performed by

Date:

23-May-02

I

EQUIPMENT: I

PAR# 5365

TEST PLAN: 288117-70020

Patient Programmer for Neuro devices.

DATE: 29 MAY 02

INITIAL VISUAL &amp; ELECTRICAL

## SUMMARY SHEET

TECH:

SERIAL#	VISUAL	Requestor did functional
NJD000109P	O.K.	X
NJD000110P	O.K.	X
NJD000111P	O.K.	X
NJD000113P	O.K.	X
NJD000114P	O.K.	X
NJD000116P	O.K.	X
NJD000119P	O.K.	X
NJD000120P	O.K.	X
NJD000121P	O.K.	X
NJD000122P	O.K.	X
NJD000123P	O.K.	X
NJD000124P	O.K.	X
NJD000126P	O.K.	X
NJD000127P	O.K.	X
NJD000128P	O.K.	X
NJD000129P	O.K.	X
NJD000130P	O.K.	X
NJD000131P	O.K.	X
NJD000133P	O.K.	X
NJD000134P	O.K.	X
NJD000136P	O.K.	X
NJD000137P	O.K.	X
	29-May	

RESULTS:

Exhibit D (cont.)

SUMMARY SHEET

PAR# 5365      TEST PLAN: 288117-70020      TECH:  
Patient Programmer for Neuro devices.      Broad Band Random Vibration paragraph 6.2.4 of test plan.  
DATE: 4-Jun-02      All Functional Testing done per 6.1 except backlight and IR port.  
Subject samples to

SERIAL#	Back down	Visual	R. side dow	Visual	Top up	Visual	Functional	Observations
NJD000109P								
NJD000110P								
NJD000111P								
NJD000113P								
NJD000114P								
NJD000116P								
NJD000119P								
NJD000120P								
NJD000121P								
NJD000122P								
NJD000123P								
NJD000124P								
NJD000126P								
NJD000127P								
NJD000128P								
NJD000129P								
NJD000130P								
NJD000131P								
NJD000133P								
NJD000134P								
NJD000136P								
NJD000137P								
Date Completed	7-Jun	7-Jun	7-Jun	7-Jun	7-Jun	7-Jun	13-Jun	

NOTES:    A=  
              B=  
              C=

RESULTS:

EQUIPMENT:



PAR# 5365

TEST PLAN: 288117-70020

TECH: ROY POPE

Patient Programmer for Neuro devices.

Mechanical Shock paragraph 6.2.5 of test plan.

DATE:

20-Jun-02

All Functional Testing done per 6.1 except backlight and IR port.

Subject samples to

SERIAL#	Front	Back	Top	Bottom	Left side	Right side	Testing
NJD000109P							
NJD000110P							
NJD000111P							
NJD000113P							
NJD000114P							
NJD000116P							
NJD000119P							
NJD000120P							
NJD000121P							
NJD000122P							
NJD000123P							
NJD000124P							
NJD000126P							
NJD000127P							
NJD000128P							
NJD000129P							
NJD000130P							
NJD000131P							
NJD000133P							
NJD000134P							
NJD000136P							
NJD000137P							

NOTES:

A=

B=

RESULTS:

EQUIPMENT:

Exhibit D (cont.)

**Test Path #3 from DVT Plan 288117-70020 Section 7.0****DVT Pre-Test Performed to verify operational units.**

Buttons

Serial Number	operational	Audio	LCD	Battery contact	Battery Door	Real time clock	IR	Backlight	Communication	Results
NJD000081P	x	x	x	x	x	x	x	x	x	OK
NJD000082P	x	x	x	x	x	x	x	x	x	OK
NJD000083P	x	x	x	x	x	x	x	x	x	OK
NJD000084P	x	x	x	x	x	x	x	x	x	OK
NJD000086P	x	x	x	x	x	x	x	x	x	OK
NJD000087P	x	x	x	x	x	x	x	x	x	OK
NJD000089P	x	x	x	x	x	x	x	x	x	OK
NJD000092P	x	x	x	x	x	x	x	x	x	OK
NJD000093P	x	x	x	x	x	x	x	x	x	OK
NJD000094P	x	x	x	x	x	x	x	x	x	OK
NJD000096P	x	x	x	x	x	x	x	x	x	OK
NJD000097P	x	x	x	x	x	x	x	x	x	OK
NJD000098P	x	x	x	x	x	x	x	x	x	OK
NJD000099P	x	x	x	x	x	x	x	x	x	OK
NJD000100P	x	x	x	x	x	x	x	x	x	OK
NJD000101P	x	x	x	x	x	x	x	x	x	OK
NJD000102P	x	x	x	x	x	x	x	x	x	OK
NJD000103P	x	x	x	x	x	x	x	x	x	OK
NJD000104P	x	x	x	x	x	x	x	x	x	OK
NJD000106P	x	x	x	x	x	x	x	x	x	OK
NJD000107P	x	x	x	x	x	x	x	x	x	OK
NJD000108P	x	x	x	x	x	x	x	x	x	OK

Exhibit D (cont.)

Date: 23-May-02

Testing performed by

EQUIPMENT: I

PAR# 5365  
 Patient Programmer for Neuro devices.  
 DATE: 29 MAY 02

TECH:

TEST PLAN: 288117-70020  
 INITIAL VISUAL & ELECTRICAL

SERIAL#	VISUAL	Requestor
NJD000081P	O.K.	did functional
NJD000082P	O.K.	X
NJD000083P	O.K.	X
NJD000084P	O.K.	X
NJD000086P	O.K.	X
NJD000087P	O.K.	X
NJD000089P	O.K.	X
NJD000092P	O.K.	X
NJD000093P	O.K.	X
NJD000094P	O.K.	X
NJD000096P	O.K.	X
NJD000097P	O.K.	X
NJD000098P	O.K.	X
NJD000099P	O.K.	X
NJD000100P	O.K.	X
NJD000101P	O.K.	X
NJD000102P	O.K.	X
NJD000103P	O.K.	X
NJD000104P	O.K.	X
NJD000106P	O.K.	X
NJD000107P	O.K.	X
NJD000108P	O.K.	X
Date: Complete	29-May	

RESULTS: NO ANOMALIES NOTED



I=

EQUIPMENT:

.....

Section 6.3.6 Button Endurance - Specification for KSS321G, used for  
... buttons (  
Life Cycle data show life expectancy of

Section 6.3.8 Flamability - Both top and bottom housings are made from

Other Data

DVT Test Data for 288117-70020

Revision 4.0



Section 6.3.6 - Button Endurance - Specification for KSC621- Used for  
top buttons (  
Life Cycle data show life expectancy



Exhibit D (cont.)

Other Data

DVT Test Data for 288117-70020

Revision 4.0


Exhibit D (cont.)

Other Data

DVT Test Data for 288117-70020

Revision 4.0

Section 6.3.6 - Button Endurance - Specification for F  
Used for  
Life Cycle data show  
life expectancy of



Blank Page

Exhibit D (cont.)

Other Data

DVT Test Data for 288117-70020

Revision 4.0

Exhibit D (cont.)

Other Data

DVT Test Data for 288117-70020

Revision 4.0

Section 6.3.7 Scratch resistance -

—



Exhibit D (cont.)

Other Data

DVT Test Data for 288117-70020

Revision 4.0



Exhibit D (cont.)

Other Data

DVT Test Data for 288117-70020

Revision 4.0